

**बिड दस्तावेज़ / Bid Document**

| बिड विवरण / Bid Details                                                                                                      |                                                                   |
|------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------|
| बिड बंद होने की तारीख/समय / Bid End Date/Time                                                                                | 17-08-2026 17:00:00                                               |
| बिड खुलने की तारीख/समय / Bid Opening Date/Time                                                                               | 17-08-2026 17:30:00                                               |
| बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)                                                  | 30 (Days)                                                         |
| मंत्रालय/राज्य का नाम / Ministry/State Name                                                                                  | Madhya Pradesh                                                    |
| विभाग का नाम / Department Name                                                                                               | Public Health And Family Welfare Department Madhya Pradesh        |
| संगठन का नाम / Organisation Name                                                                                             | N/a                                                               |
| कार्यालय का नाम / Office Name                                                                                                | Cmho Badwani                                                      |
| शिकायत निवारण के संपर्क विवरण / Contact details of Grievance redressal                                                       | HOD Email id :cmhobar@nic.in<br>Buyer Email id: dpmubar-mp@nic.in |
| कुल मात्रा / Total Quantity                                                                                                  | 52                                                                |
| वस्तु श्रेणी / Item Category                                                                                                 | Infant Warmer (V2) (Q2)                                           |
| बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)              | 13 Lakh (s)                                                       |
| मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) / OEM Average Turnover (Last 3 Years)                                     | 110 Lakh (s)                                                      |
| उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service             | 3 Year (s)                                                        |
| टर्नओवर पात्रता मानदंड / Turnover eligibility criteria                                                                       | To be verified by the buyer at the time of technical evaluation   |
| वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover                 | Yes   Complete                                                    |
| स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years Of Experience and Turnover | Yes   Complete                                                    |

| बिड विवरण/Bid Details                                                                                                                                                                                                    |                                                                                                                                                                                                                                                                                                                                                                         |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| विक्रेता से मांगे गए दस्तावेज़/Document required from seller                                                                                                                                                             | Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC)<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid? | Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)                                                                                                                                                                                                  |
| बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension                                                                                | 3                                                                                                                                                                                                                                                                                                                                                                       |
| दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended                                                                                                      | 3                                                                                                                                                                                                                                                                                                                                                                       |
| ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count                                                                                                                                            | 1                                                                                                                                                                                                                                                                                                                                                                       |
| विगत प्रदर्शन /Past Performance                                                                                                                                                                                          | 50 %                                                                                                                                                                                                                                                                                                                                                                    |
| बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled                                                                                                                                                                       | No                                                                                                                                                                                                                                                                                                                                                                      |
| बिड का प्रकार/Type of Bid                                                                                                                                                                                                | Two Packet Bid                                                                                                                                                                                                                                                                                                                                                          |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation                                                                                        | 2 Days                                                                                                                                                                                                                                                                                                                                                                  |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)                                     | No                                                                                                                                                                                                                                                                                                                                                                      |
| मूल्यांकन पद्धति/Evaluation Method                                                                                                                                                                                       | Total value wise evaluation                                                                                                                                                                                                                                                                                                                                             |
| मध्यस्थता खंड/Arbitration Clause                                                                                                                                                                                         | No                                                                                                                                                                                                                                                                                                                                                                      |
| सुलह खंड/Mediation Clause                                                                                                                                                                                                | No                                                                                                                                                                                                                                                                                                                                                                      |

#### ईएमडी विवरण/EMD Detail

|                             |                     |
|-----------------------------|---------------------|
| एडवाइजरी बैंक/Advisory Bank | State Bank of India |
| ईएमडी राशि/EMD Amount       | 83000               |

#### ईपीबीजी विवरण /ePBG Detail

|                   |    |
|-------------------|----|
| आवश्यकता/Required | No |
|-------------------|----|

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

#### लाभार्थी /Beneficiary :

Store Keeper

Cmho Badwani, Public Health and Family Welfare Department Madhya Pradesh, N/A,

(Sarjan Singh Bheel)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमएसई खरीद वरीयता/MSE Purchase Preference

|                                           |    |
|-------------------------------------------|----|
| एमएसई खरीद वरीयता/MSE Purchase Preference | No |
|-------------------------------------------|----|

#### एमआईआई खरीद वरीयता/MII Purchase Preference

|                                            |    |
|--------------------------------------------|----|
| एमआईआई खरीद वरीयता/MII Purchase Preference | No |
|--------------------------------------------|----|

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.

2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.

3. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.

4. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.

5. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

6. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

7. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

## Infant Warmer (V2) ( 52 pieces )

### तकनीकी विशिष्टियाँ /Technical Specifications

[\\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

| विवरण/Specification | विशिष्टि का नाम /Specification Name                                                                  | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)                                                                                                                                                                                                                                          |
|---------------------|------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| GENERAL             | Product Description                                                                                  | Infant Radiant Warmer                                                                                                                                                                                                                                                                                    |
| PRODUCT INFORMATION | Facility to display set and observed skin temperature in degree C and heater power output separately | Yes                                                                                                                                                                                                                                                                                                      |
|                     | User friendly touch panel control                                                                    | Yes                                                                                                                                                                                                                                                                                                      |
|                     | Type of heater                                                                                       | Ceramic / High Alloy Steel / Quartz infrared / Calrod heater                                                                                                                                                                                                                                             |
|                     | Audiovisual alarm facility                                                                           | Overheating beyond set temperature range, Spurt in Heater output i.e. more than 60% heater output for 10 minutes system should raise alarm and cutoff heater, Power failure, Heater failure, Patient temperature less/Greater than required/set temperature, Probe failure, Temperature Probe disconnect |
|                     | Warmer head should be rotatable in different direction, so as to allow taking X-ray                  | Yes                                                                                                                                                                                                                                                                                                      |
|                     | Observation light Luminance                                                                          | ? 1000 Lux                                                                                                                                                                                                                                                                                               |
|                     | Colour temperature                                                                                   | 3700K to 5100K                                                                                                                                                                                                                                                                                           |
|                     | Setable Temperature                                                                                  | 32 to 38 Degree C                                                                                                                                                                                                                                                                                        |
|                     | Temperature resolution                                                                               | ? 0.1 Degree C                                                                                                                                                                                                                                                                                           |
|                     | Temperature accuracy                                                                                 | ? 0.2 Degree C                                                                                                                                                                                                                                                                                           |

| विवरण/Specification | विशिष्टि का नाम /Specification Name                                                                                                                                      | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|
|                     | Facility to lock the keyboard to avoid unwanted user modification of the set parameters                                                                                  | Yes                                                             |
|                     | Should have a Feather Touch operation with large digital display and comprehensive alarms, Control Panel should be liquid proof and allow easy and hygienic disinfection | Yes                                                             |
|                     | Manual Mode can adjust Heater Output 10 -100 %, with 10% increment                                                                                                       | Yes                                                             |
|                     | In Manual Mode an auditory and visual alarm shall be given at least every 15 minutes                                                                                     | Yes                                                             |
|                     | Bed height from Floor (in cm)                                                                                                                                            | 80 to 100                                                       |
|                     | <b>Height Adjustable feature</b>                                                                                                                                         | Manual                                                          |
|                     | Bed Distance from the heat source (in cm)                                                                                                                                | 80 to 90                                                        |
|                     | Lockable castor wheels                                                                                                                                                   | Yes                                                             |
|                     | Power ON Indication                                                                                                                                                      | Yes                                                             |
|                     | Provision for X-Ray cassette holder                                                                                                                                      | Yes                                                             |
|                     | Markings on the bassinet and X-Ray cassette holder to enable proper positioning of the baby while doing the X-Ray                                                        | Yes                                                             |
|                     | Number of tubing ports on the side walls (For cable management)                                                                                                          | 2 to 4                                                          |
|                     | Height of the side walls over the mattress (in mm)                                                                                                                       | ? 110                                                           |
|                     | Bay bed should be crevice free for ease of cleaning, infection control                                                                                                   | Yes                                                             |
|                     | Mattress used should be of biocompatible material                                                                                                                        | Yes                                                             |
|                     | Transparent collapsible side walls easily detachable for cleaning                                                                                                        | Yes                                                             |
|                     | Skin temperature probe with biocompatible baby contact material                                                                                                          | Yes                                                             |
|                     | Mobility, portability on castors                                                                                                                                         | Yes                                                             |

| विवरण/Specification      | विशिष्टि का नाम /Specification Name                                                                                                | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------|
|                          | 2/4 of the castors should have brakes                                                                                              | Yes                                                                                                  |
|                          | Castor size                                                                                                                        | ? 4 inch                                                                                             |
|                          | Tilt-able bed (10 to 12 Deg ) Lockable in position at any angle                                                                    | Yes                                                                                                  |
| Settings                 | <b>Settings with facility to clearly display the selected mode</b>                                                                 | Manual mode and Servo mode                                                                           |
|                          | Provision for Set temperature range (in servo mode)                                                                                | 32 to 38 Degree C                                                                                    |
|                          | User's interface                                                                                                                   | Manual and Servo controlled temperature regulation                                                   |
|                          | Display                                                                                                                            | LCD/ LED                                                                                             |
|                          | Software and/or standard of communication(where ever required)                                                                     | Inbuilt software; Interruption and restoration of the power supply does not change the preset values |
|                          | Device shall not overbalance when placed in any transport position of normal use on a 10° inclined plane from the horizontal plane | Yes                                                                                                  |
|                          | Transformers of devcie shall be protected against overheating in the event of short circuit or overload of any output winding      | Yes                                                                                                  |
|                          | Temperature on the baby mattress when the warmer is operating under steady temperature condition                                   | Not exceed 43 Degree C                                                                               |
| PHYSICAL CHARACTERISTICS | Temperature differences on the mattress                                                                                            | Not exceed 2 Degree C                                                                                |
|                          | Overall Weight of equipment                                                                                                        | ? 50 Kgs                                                                                             |
|                          | Mattress Dimension Length                                                                                                          | 60 Cms or more                                                                                       |
|                          | Mattress Dimension Width                                                                                                           | 40Cms or more                                                                                        |
|                          | Auditory alarm shall be audible with minimum distance of 6 feet                                                                    | Yes                                                                                                  |
|                          | Should maintain desired set temperature and uniform heat distribution in baby cradle                                               | Yes                                                                                                  |
| ACCESSORIES              | Number of IV pole                                                                                                                  | 1                                                                                                    |
|                          | Number of Storage drawer                                                                                                           | 2, Not Provided                                                                                      |
|                          | Number of Accessory tray (Minimum weight bearing capacity of 5Kg or more)                                                          | 1                                                                                                    |

| विवरण/Specification    | विशिष्टि का नाम /Specification Name                                                                                                                                                        | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------|
|                        | Number of X- ray tray                                                                                                                                                                      | 1                                                               |
|                        | Number of Mattress                                                                                                                                                                         | 1                                                               |
| ELECTRICAL REQUIREMENT | Power Requirements                                                                                                                                                                         | 220 to 240V, 50 Hz AC, $\pm 10\%$ of input                      |
| CERTIFICATIONS         | Compliance to Medical Device Rules (MDR) 2017 as amended till date                                                                                                                         | Yes                                                             |
|                        | Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date | Yes                                                             |
|                        | Certification for manufacturing unit                                                                                                                                                       | ISO:13485 (Latest)                                              |
|                        | Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date                                                                     | Yes                                                             |
|                        | Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement                           | Yes                                                             |
|                        | Electrical Safety Compliance Standard                                                                                                                                                      | IEC 60601-1-2 or Equivalent BIS                                 |
| WARRANTY               | <b>Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b>                                       | 3, 4, 5 Or higher (year)                                        |

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

| क्र.सं./S.No. | परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer | पता/Address                                                | मात्रा /Quantity | डिलीवरी के दिन/Delivery Days |
|---------------|---------------------------------------------------------|------------------------------------------------------------|------------------|------------------------------|
| 1             | Sarjan Singh Bheel                                      | 451551,District Drug Store CMHO, TB Ward Compound, Barwani | 52               | 15                           |

## Special terms and conditions-Version:1 effective from 01-08-2024 for category Infant Warmer (V2)

1. 1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of

Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.

12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

### 2. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

**1.NOTE:** The bidder must have an annual turnover of not less than INR **13 lakh** in each of the last three financial years.

**2.Sample Clause:** After award of contract successful bidder shall have to get advance sample approval from buyer before supply.

3.The documents and specifications uploaded under ATC shall form an integral part of the bid and shall be mandatory for all bidders.

#### 3.Servo Controlled Radiant Warmer Specifications:

**1.** It should be microprocessor-based servo-controlled unit that will allow at least two operation modes- servo (automatic) and manual mode.

**2.** The radiant warmer should have a heating element with power consumption of 800-1200 Watts.

**3.** It should have quartz / non-ceramic infrared / Ceramic Heater heating element with parabolic stainless-steel reflector for uniform radiation of heat waves to all corners of the basinet.

**4.** Heater should be protected by a metal grid to prevent shock hazards.

**5.** The heater box/head should be either of swivel type and it must rotate to  $\pm 90^\circ$  from its axis or fixed type head provided it allows X-Ray exposure from top.

**6.** It should have LED display (at least 1") for showing skin temperature in manual mode and set and skin temperatures in servo mode. Temperature display should be bright, numerical.

**7.** The unit may have a rechargeable battery backup (30 min or more) for display of temperatures & alarm during power failure.

**8.** It should have an adjustable light source (LED) for observation of the baby.

**9.** Probe: Skin Probe and Air Probe/Sensor.

**10.** Should have facility to lock the Keyboard to avoid unwanted user modification of the set parameters.

**11. The basinet:**

**i.** Comprises of foldable/dropdown transparent acrylic/polycarbonate side panels of minimum 6 mm in thickness and can be swiveled or collapsible easily for cleaning and procedures.

**Note:** There should be provision for under surface phototherapy unit placement in the basinet of Servo Controlled Radiant Warmer.

**ii.** Should not be fixed with the warmer; if it is fixed then it should be detachable easily for cleaning purposes. Both should move on lockable wheels.

**iii.** It must have one fixed transparent acrylic sheet below mattress for under surface phototherapy and an additional acrylic tray with basinet for positioning the X-ray cassette.

**iv.** The unit should have facility for inclination in both directions (Trendelenburg and reverse Trendelenburg) minimum up to 12°.

**v.** Stitch-less mattress should be supplied with a foam density of 25 kg/m<sup>3</sup> with at least 1 inch in thickness.

**vi.** Mattress cover should be made of fire-retardant material.

**vii.** Mattress should be waterproof and easily washable. It should be resistant to cleaning with chlorine-based solution.

**12. Dimensions:**

**i.** Probe length: Not less than 3 ft. (Skin and Air Probe).

**ii.** Length of basinet: Minimum 680 mm. Width of Basinet: Minimum 500 mm.

**iii.** Working level from the ground up to upper level of the mattress: Minimum 870 mm.

**iv.** Distance from mattress top and heat source: Minimum 850 mm.

**v.** Diameter of castors for Radiant Warmer and Basinet with brakes/locks on front wheels: Minimum 75 mm.

**vi.** The unit should be supplied with 2 (two) drawers and 1 (one) IV stand. The minimum diameter of IV stand (stainless steel) should be 2.5 cm with 4 pegs (stainless steel) for hanging IV fluids.

**vii.** Height adjustment system for Radiant Warmer and Basinet should be accommodated: ±2 inch.

**viii.** Monitor tray (to be supplied with): 300 mm × 270 mm or more.

**13.** Operating temperature range: 10°C to 30°C or wider.

**14.** Operating humidity range: 30%–90% RH.

**15.** Skin temperature:

- Range: 32°C to 38°C
- Accuracy: ±0.2°C
- Resolution: 0.1°C

**16.** Alarm for Probe failure, Heat failure, High Temperature, Low Temperature and Power failure.

**17.** Electrical supply: Power input should be 180–240 VAC, 50 Hz, fitted with Indian 3-pin plug.

**18.** UPS of suitable rating with voltage regulation and spike protection for 60 minutes backup.

**19.** Accessories & Consumables to be supplied with the machine:

**i.** Mattress: 02 (Two)

**ii.** Temperature (Skin) Probe and Air Probe (Reusable)/Air Sensor: 02 Each (Two)

**iii.** Fuse Set: 01 (One)

iv. Monitor Tray: 01 (One)

20. Documents/Certification:

i. It should be USFDA / European CE (4-digit Notified Body) / BIS approved.

ii. CDSCO License/Registration.

iii. Manufacturer should be ISO 13485 Certified.

iv. IEC 60601 Test Report.

21. User/Technical/Maintenance Manual to be supplied in English/Hindi.

22. Cleaning and maintenance of Radiant Warmer and Basinet should be user friendly.

23. Warranty: 03 Years

24. CMC: 05 Years

.

3. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer, is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1

bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.

14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

**This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.**

**However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**