

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	26-09-2026 11:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	27-09-2026 11:00:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	95 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Steel
विभाग का नाम/Department Name	Steel Authority Of India Limited
संगठन का नाम/Organisation Name	Chandrapur Ferro Alloy Plant
कार्यालय का नाम/Office Name	Chandrapur
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :dvrc3@sail.in Buyer Email id: radha.sahni@mjunction.in
कुल मात्रा/Total Quantity	7
वस्तु श्रेणी /Item Category	Breath Analyser (Q3)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	No
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3

बिड विवरण/Bid Details	
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	Breath Analyser
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Payment Timelines	Payments shall be made to the Seller within 30 days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

आवश्यकता/Required	No
-------------------	----

ईपीबीजी विवरण /ePBG Detail

आवश्यकता/Required	No
-------------------	----

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
---	-----

सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ / Purchase Preference to MSE OEMs/ Service Provider available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

ट्रेड्स भुगतान संबंधी विवरण/TReDS Payment Details

This Bid provides for Trade Receivables Discounting System (TReDS) as Preferred mode of payment. For MSME sellers, payments may be processed through a TReDS exchange in which the Buyer is registered, subject to applicable policy and regulatory guidelines. Accordingly, sellers intending to avail payment through TReDS are required to be registered with at least one TReDS exchange in which the buyer is registered.

S.No.	Name
1	RXIL
2	M1XCHANGE
3	INVOICEMART

1. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted in case of product bids, whereas in case of services the eligibility is automatically validated. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within $L-1+ 15\%$ (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

2. Short Duration Bid has been published by the Buyer with the approval of the Competent authority due to Emergency procurement of critical products/services.

Breath Analyser (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Generic	Sensor Type used for measurement	Electrochemical Fuel Cell

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Type of Analyser	Stationary (Wall Mounted or Table Top)
	Sampling Method	Non-Contact
Performance	Detection Range	upto 5.00 mg/L
	Memory Capacity for Test Record (Minimum)	55000-60000
	Response Time (in Second)	2.5-5
	Reset Time (in Second)	5-10

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Niraj Prakash Tembhekar	442401,SAIL-Chandrapur Ferro Alloy Plant, Mul Road, Chandrapur	1	60

Breath Analyser (6 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Generic	Sensor Type used for measurement	Electrochemical Fuel Cell
	Type of Analyser	Handheld
	Sampling Method	Both (Contact and Non Contact)
	Alarm System Type	Visual (LED Indication), Audio
	Data Transfer Options	USB
Performance	Detection Range	upto 5.00 mg/L
	Memory Capacity for Test Record (Minimum)	50000-55000
	Response Time (in Second)	2.5-5
	Reset Time (in Second)	10-15

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Minimum Operating Temperature Range (degree Celsius)	-5
Inbuilt Display	Display Type	LED
Features	Availability of Inbuilt Printer	Yes
	Low Battery Indication	Yes
Power Source	Power Source	Battery
	Battery Type	Rechargeable
	Maximum Battery Backup time (in Hours)	7-9

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Niraj Prakash Tembhekar	442401,SAIL-Chandrapur Ferro Alloy Plant, Mul Road, Chandrapur	6	60

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

MATERIAL DESCRIPTION:

Sr. No.1

Full description: Stationary Breath Analyser

Stationary Wall Mounted / Table Top Breath Analyser with Electrochemical Fuel Cell Sensor, having the following minimum specifications:

Sensor Type: Electrochemical Fuel Cell.

Type of Analyser: Stationary – Wall Mounted / Table Top.

Sampling Method: Non-Contact.

Detection Range: Up to 5.00 mg/L.

Accuracy: $\pm 0.005\%$ BAC at 0.050% BAC or better.

Response Time: 2.5–5 seconds or better.

Reset Time: 5–10 seconds or better.

Memory Capacity: Minimum 55,000 test records.

Measurement Units: mg/L, mg/100 ml, %BAC and g/L.

Alarm System: Visual LED indication.

Data Transfer: USB.

Display: Inbuilt LCD display.

Protection: Minimum IP54.

Operating Temperature: -10°C to $+55^{\circ}\text{C}$ or wider.

Inbuilt Printer: Yes.

Power Supply: 230 V, 50 Hz AC.

The equipment shall be supplied complete with suitable power adaptor, mounting arrangement and all standard accessories required for satisfactory installation and operation.

Quantity: 1 EACH

Sr. No.2

Full description:

Handheld Breath Analyser

Portable Handheld Breath Analyser with Electrochemical Fuel Cell Sensor, having the following minimum specifications:

Sensor Type: Electrochemical Fuel Cell.

Type of Analyser: Handheld / Portable.

Sampling Method: Both Contact and Non-Contact.

Detection Range: Up to 5.00 mg/L.

Accuracy: $\pm 0.005\%$ BAC at 0.050% BAC or better.

Response Time: 2.5–5 seconds or better.

Reset Time: 10–15 seconds or better.

Memory Capacity: Minimum 50,000 test records.

Alarm System: Visual LED indication and Audio Alarm.

Data Transfer: USB.

Display: Inbuilt display of minimum 2 inch.

Protection: Minimum IP65.

Operating Temperature: -5°C to +50°C or wider.

Inbuilt Printer: Yes.

Power Source: Rechargeable Battery.

Battery Backup: Minimum 7 hours.

Low Battery Indication: Yes.

The equipment shall be supplied complete with charger/adaptor, carry case, mouthpiece, disposable blowing nozzle and all standard accessories required for satisfactory operation.

Quantity: 6 EACH

Eligibility Criteria:

1. The bidder/OEM should have successfully supplied same or similar Breath Analysers / Alcohol Breath Analysers / Alcohol Detection Equipment during the preceding five (05) years to any Central/State Government Department, PSU, Government Undertaking, Autonomous Body, reputed industrial establishment or similar institution.
2. Documentary evidence such as Purchase Order / GeM Contract / Work Order along with satisfactory supply/completion credentials may be sought for verification.
3. The offered equipment shall conform to the detailed technical specifications stipulated in the PR.

Special Terms & Conditions:

1. SCOPE OF SUPPLY / WORK

The scope shall include supply, installation, commissioning, testing and demonstration of Breath Analysers at designated locations of SAIL-CFP, complete with all required accessories, power adaptor/charger, mounting arrangements, printer, sampling accessories, cables and other items necessary for satisfactory operation.

The supplier shall provide initial calibration, functional testing and basic operational demonstration/training to the concerned personnel. The equipment shall be handed over in fully functional and ready-to-use condition.

For Stationary Breath Analysers, installation shall include wall/table-top mounting and connection at the designated location, as applicable. For Handheld Breath Analysers, commissioning shall include functional testing of the analyser, battery/charger, printer, alarm and sampling modes.

2. DETAILED TECHNICAL SPECIFICATION

Item 1 – Stationary Breath Analyser

Stationary Wall Mounted / Table Top Breath Analyser with Electrochemical Fuel Cell Sensor, having the following minimum specifications:

Sensor Type: Electrochemical Fuel Cell.

Type of Analyser: Stationary – Wall Mounted / Table Top.

Sampling Method: Non-Contact.

Detection Range: Up to 5.00 mg/L.

Accuracy: $\pm 0.005\%$ BAC at 0.050% BAC or better.

Response Time: 2.5–5 seconds or better.

Reset Time: 5–10 seconds or better.

Memory Capacity: Minimum 55,000 test records.

Measurement Units: mg/L, mg/100 ml, %BAC and g/L.

Alarm System: Visual LED indication.

Data Transfer: USB.

Display: Inbuilt LCD display.

Protection: Minimum IP54.

Operating Temperature: -10°C to $+55^{\circ}\text{C}$ or wider.

Inbuilt Printer: Yes.

Power Supply: 230 V, 50 Hz AC.

The equipment shall be supplied complete with suitable power adaptor, mounting arrangement and all standard accessories required for satisfactory installation and operation.

The above requirements are based on the applicable GeM product specifications for the stationary Electrochemical Fuel Cell type Breath Analyser.

Item 2 – Handheld Breath Analyser

Portable Handheld Breath Analyser with Electrochemical Fuel Cell Sensor, having the following minimum specifications:

Sensor Type: Electrochemical Fuel Cell.

Type of Analyser: Handheld / Portable.

Sampling Method: Both Contact and Non-Contact.

Detection Range: Up to 5.00 mg/L.

Accuracy: $\pm 0.005\%$ BAC at 0.050% BAC or better.

Response Time: 2.5–5 seconds or better.

Reset Time: 10–15 seconds or better.

Memory Capacity: Minimum 50,000 test records.

Alarm System: Visual LED indication and Audio Alarm.

Data Transfer: USB.

Display: Inbuilt display of minimum 2 inch.

Protection: Minimum IP65.

Operating Temperature: -5°C to $+50^{\circ}\text{C}$ or wider.

Inbuilt Printer: Yes.

Power Source: Rechargeable Battery.

Battery Backup: Minimum 7 hours.

Low Battery Indication: Yes.

The equipment shall be supplied complete with charger/adaptor, carry case, mouthpiece, disposable blowing nozzle and all standard accessories required for satisfactory operation.

The above requirements are based on the applicable GeM product specifications for the handheld Electrochemical Fuel Cell type Breath Analyser.

3. DOCUMENTS TO BE SUBMITTED

The supplier shall submit the following documents along with the equipment, wherever applicable:

Invoice / Tax Invoice. OEM Technical Data Sheet / Product Catalogue for the supplied make and model.

Valid Calibration Certificate for each instrument indicating instrument serial number.

Test Certificate / Conformity Certificate, wherever applicable. Guarantee / Warranty Certificate.

Make, Model and Serial Number details.

OEM User / Operating / Calibration / Maintenance Instructions.

Recommended calibration interval/frequency.

OEM Authorization / Dealership / Distributorship Certificate, where the supplier is not the OEM, wherever applicable.

List of accessories supplied.

Packing List / Delivery Challan indicating item-wise quantity.

Transporter LR Copy.

All applicable certificates and documents shall be submitted along with the equipment and shall form part of inspection and acceptance at SAIL-CFP, Chandrapur.

4. GUARANTEE / WARRANTY

The complete equipment shall carry a minimum warranty of 01 (One) year from the date of successful installation, commissioning and acceptance at SAIL-CFP, Chandrapur.

The warranty shall cover manufacturing defects and functional defects in the sensor, display, printer, alarm, electronics, battery/charging system, power supply/adaptor and other components supplied with the equipment.

Any equipment/component found defective due to manufacturing defect or defective workmanship during the warranty period shall be repaired or replaced by the supplier without any additional cost to SAIL-CFP.

5. CALIBRATION

Each Breath Analyser shall be supplied duly calibrated and ready for use with a valid calibration certificate traceable to the respective instrument serial number.

The supplier/OEM shall clearly indicate the recommended calibration interval/frequency and provide details of the facility/support available for subsequent calibration.

6. INSTALLATION, COMMISSIONING and DEMONSTRATION

The supplier shall carry out installation, commissioning and functional testing of the equipment at locations designated by SAIL-CFP.

After commissioning, the supplier shall demonstrate the operation of the equipment including sampling/testing procedure, interpretation of results, alarm operation, printing, data transfer, charging/power operation, routine checks and basic maintenance, as applicable.

Installation, commissioning, testing and demonstration shall be completed without any additional cost to SAIL-CFP.

7. INSPECTION and ACCEPTANCE

The equipment shall be inspected at SAIL-CFP for quantity, physical condition, make/model, serial number and conformity with the ordered technical specifications.

Functional checking of alcohol detection, sampling modes, display, alarm, printer, USB/data-transfer facility, battery/charger, power supply and other applicable functions shall be carried out.

Accessories, calibration certificates, warranty certificates, OEM technical documents and other specified documents shall also be verified.

Final acceptance shall be subject to successful installation, commissioning, functional testing and demonstration of the complete equipment.

8. DELIVERY SCHEDULE

The complete scope covering supply, installation, commissioning, testing and demonstration shall be completed within 02 (Two) months from the date of Purchase Order / GeM Contract.

Delivery Location: Stores, SAIL – Chandrapur Ferro Alloy Plant (CFP), Chandrapur, Maharashtra.

Installation and commissioning shall be carried out at the designated locations within SAIL-CFP as advised by the Indenting Department.

9. PAYMENT TERMS

Payment shall be made as per applicable GeM terms and conditions after completion of supply, installation, commissioning, successful functional testing, demonstration and final acceptance of the complete equipment at SAIL-CFP, Chandrapur.

Spl.Del/Pkg/Insp Instructions:

The equipments shall be inspected at SAIL-CFP for quantity, physical condition, make/model, serial number, technical specification, functionality, accessories, calibration certificate and warranty documents. Functional checking of detection, display, alarm, printer, USB/data transfer and battery/power supply, as applicable, may be carried out during inspection.

General terms & Condition:

The tenderer will be responsible for building awareness among their concerned employees about the applicable rules regarding environment.

All other terms & conditions as per SAIL P1.

Anti-bribery Undertaking: By participating in this tender, the suppliers / Vendors/Contractors are deemed to have undertaken that they shall not give or take any financial or non-financial bribe, to or from anyone d

uring the tender or during the execution of the contract thereafter and if they notice any such incident hap
pening , they shall report it to SAIL Vigilance.

Instruction for bid submission:

Bidders are instructed to submit documents sought in the bid at the time of bid submission. Non submissio
n of requisite documents may lead to rejection of the bid without any issue of clarification. Rejected vendo
rs shall be formally communicated with reasons for rejection and provided a minimum period of two (2) wo
rking days to submit representations limited strictly to clarifications and supporting documents related to d
ocuments already submitted.

All such submissions shall be evaluated, subject to consideration of only those documents existing prior to
the tender due date.

Vendors remaining ineligible shall be informed of continued rejection without any further right of represent
ation. Only techno-commercially acceptable vendors shall be considered for price discovery.

Only relevant documents to the tender are to be submitted the tender including eligibility documents, ann
exures, etc.

.

3. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/**Disclaimer**

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their
Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its
outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs
and including the modification of technical specifications and / or terms and conditions governing the bid.All
representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly
and not with GeM.If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid &
resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel
the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such
Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of
Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such
suspension period has already expired.
4. Mandating submission of documents (other than EMD and signed Integrity Pact - which can be submitted
within 5 days of bid opening) in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as
the case may be, shall be permitted in cases where trial / sample are allowed as per approved and
published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector
Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy
of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be

determined and redressed by the concerned Buyer Organisation only.

8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and
- All operative provisions of the erstwhile Labour Laws until their complete substitution.

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a

land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---