बिड संख्या/Bid Number: GEM/2026/B/8243451
दिनांक /Dated: 05-11-2026

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	20-11-2026 14:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	21-11-2026 14:00:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Andaman & Nicobar
विभाग का नाम/Department Name	Health Department Andaman & Nicobar
संगठन का नाम/Organisation Name	N/a
कार्यालय का नाम/Office Name	Aniims
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :diraniims.and@nic.in Buyer Email id: buycon59.hdan.an@gembuyer.in
कुल मात्रा/Total Quantity	2
वस्तु श्रेणी /Item Category	Anaesthesia Workstation (V3) (Q2)
पिछले 3 वितीय वर्षों का बिडर का न्यूनतम औसत वार्षिक वित्तीय टर्नओवर/Minimum Average Annual Turnover of the bidder (For the last 3 Financial Years)	62 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	62 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
टर्नओवर पात्रता मानदंड/Turnover eligibility criteria	To be verified by the buyer at the time of technical evaluation
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No

बिड विवरण/Bid Details	
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	3
विगत प्रदर्शन /Past Performance	30 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	50% Lowest Priced Technically Qualified Bidders
व्यापक रखरखाव शुल्क आवश्यक / Comprehensive Maintenance Charges Required	Yes
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाईजरी बैंक/Advisory Bank	State Bank of India

ईएमडी राशि/EMD Amount	408000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%)/ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	62

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

लाभार्थी /Beneficiary :
Director ANIIMS
ANIIMS, DHS Annexe, Old RBD Building, Atlanta Point, Sri Vijaya Puram, South Andaman, Andaman & Nicobar Islands, PIN - 744101
(Director Aniims)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	No

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity

during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
5. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

6. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

Anaesthesia Workstation (V3) (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Components of Anesthesia workstation	Anesthesia machine, Inbuilt ventilator, Vaporizer, Gas delivery system, Multi para monitor
Gas Delivery System	Availability of Pin index yokes for gases	Oxygen and Nitrous oxide
	Number of Yokes, Oxygen	1
	Number of Yokes, N2O	1
	NIST /DISS Inlet pipeline with Hoses for O2, N2O and Air	Yes
	Availability of flow meters for O2, N2O and air	Yes
	Type of Flowmeter	Electronic
	Mode of gas mixing	Electronic

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Availability of pressure gauges for cylinder and pipelines in anesthesia workstation	Yes
	Availability of audible and visual oxygen failure alarms	Yes
	Minimum emergency oxygen flow (in liter per minute)	35 - 70
	Type of hypoxic guard with automatic cutoff of N2O	Electronic
	Minimum 25% O2 across all Oxygen - Nitrous Oxide Mixture	Yes
	Auxillary Oxygen Flowmeter	Yes
	Auxillary Common Gas outlet for use in open circuit	Yes
Vaporizer	Number of quick mount type vaporizer for easy interchangeability at a time	2
	Number of selectatec back bar	1
	Facility of interlock for vaporizer	Yes
	Minimum agent capacity for free volatile anesthetic agent (in ml)	225 ml
	Sensor connection shall be internal to prevent disconnection	Yes
	Vaporizer compatible to gases	Sevoflurane, Isoflurane, Desflurane, Halothane
	The vaporizer design should be maintenance free	Yes
	Vaporizer calibration requirement	After every 5 years
	Should have a drug specific Filling arrangement	Yes
Breathing System	Breathing system should have fresh gas de-coupled / compensation , closed circle absorber system	Yes
	Range of pressure relief valve (in mbar) of breathing system	5 to 75
	Number of steps in change over from Mechanical / Spontaneous to bag ventilation	Single step

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Number of reusable autoclavable closed silicone circuit supplied	2 Nos for adult, 2 Nos for pediatric
	Number of Disposable closed circuits - Adult	NA
	Number of Disposable closed circuits - Pediatric	NA
	Number of Disposable closed circuits - Neonatal	NA
	Number of bains circuit (adult) supplied with workstation	5
	Number of Jackson Rees circuit for pediatric patients supplied with workstation	5
	Volume capacity of breathing system (in liters)	1.5 or more
	Reusable flow sensors covered in warranty	No
	Dual flow sensor at inspiratory and expiratory port	Yes
	Breathing system should have leak and compliance test	Yes
	Breathing system should be autoclavable	Yes
	Facility of CO2 Bypass	Yes
Anesthesia Ventilator	**Available modes of operating ventilator**	Manual/spontaneous, Volume controlled, Pressure controlled, SIMV/PCV, SIMV/VCV, CMV
	Ventilator should be capable to pressure support	Yes
	Ventilator should be capable to dual mode of ventilation	Yes
	Integrated ventilator not require bellows to be change for adults and infants	Yes
	Tidal volume of ventilator (in ml)	10 to 1400 ml
	PEEP of ventilator in mbar	0 to 20 mbar
	Breathing frequency in BPM	4 to 60 BPM
	Inspiratory pause percentage of Ti	0 - 50%
	Screen size of Ventilator monitor in inches	10.1 to 15, 15 .1 and more Or higher

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Integrated ventilator shall be able to provide facility to monitor FiO2, dual agent, CO2, Automatic identification of agents, Minimal Alveolar Concentration (MAC) Value	Yes
	Integrated ventilator shall be able to provide facility to monitor expired O2 (EtO2)	Yes
	Flow Trigger Sensitivity	Adult, Pediatric, Neonatal
Monitor	Type of monitor LCD/LED/TFT	Yes
	Size of monitor in inches	15.1 to 18, 18.1 and more Or higher
	Monitor should not require lengthy start up procedure or calibration for display of parameters	Yes
	The graphical and numerical trend with split screen duration	24 hours
	Patient multipara monitor shall be able to display parameters	HR, SpO2, NIBP, ECG, Temperature, NA(If monitor not provided)
	Capnography to be displayed	No
	IBP1 and IBP2 to be displayed	Yes
	CO (cardiac output) to be displayed	Yes
	Anesthesia Gas Monitoring (AGM)	No
	BIS/ENTROPY to be displayed	Yes
	Neuro Muscular Transmission(NMT) to be displayed	Yes
	ECG monitoring with standard lead configuration, I,II,III, avL, avF&V filters so that it is protected against defibrillator, electro surgery potentials, arrhythmia analysis and ST segment analysis with trends	Yes
	NIBP should be determine with oscillometric method, should read systolic, diastolic and mean arterial pressure and pulse rate	Yes
	Pulse oximetery with sensors suitable for adults, pediatric & neonate	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	ETCO2 measurement should be by infra –red absorption technique,measuring mode should be side streamor main stream and response time should be less than 3 seconds	Yes
	Monitor should be able to measure and display all parameters in single screen	Yes
	Monitor should be provided from the same OEM	Yes
Alarms	Alarms	Tidal Volume, Minute Volume, Airway Pressure Low and High, Inspiratory Oxygen concentration, Audio power supply fail alarm, Fail to cycle warning (apnea).
	Alarms shall be adjustable to high / low limits with audio and visual alarms	Yes
Power supply	Length of main cable in meter with power socket of standard indian make to be provided	?5
	Inbuilt battery backup for monitor in Mins	60, 120, 180, 240, 300, 360 Or higher
	Power supply	220-240V AC, 50 Hz single phase
Additional Accessories	Transparent re-usable silicon face mask size 0,1, 2, 3, 4 & 5	2 each
	Breathing bag capacity - 2 Each	500 ml, 1 Litre, 2 Litre
	3/5 ECG lead	2
	Nasopharyngeal/rectal and skin Temperature probes should be Provided	2 Nos Adult, 2 Nos Pediatric, 2 Nos Neonatal
	NIBP Reusable Large adult, Reusable small adult cuffs, Reusable Pediatric cuff	2 each
	ETCO2 sample line	50
	Pressure regulator with 3 meter high Pressure tubing(1 for oxygen and 1 for Nitrous oxide)	No
	Provision for laryngscopes	Not Provided
	Number of Reusable patient circuits	2 Nos Adult, 2 Nos Pediatric
	SpO? sensor with cable provided	2 Nos Adult, 2 Nos Pediatric, 2 Nos Neonatal
	Reusable IBP interface cables	2
	Disposable IBP transducers	10

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	BIS sensors	25
	Accessories for wall mounting	NA (If trolley mounting)
ADDITIONAL PARAMETERS	Type of mounting for the Anesthesia workstation	Trolley Mounting
	Material used for Anesthesia workstation	Epoxy powder painted steel
	Mode of integrated scavenging system with pressure relief valve	Active
	Minimum Battery backup for anesthesia workstation	1 hour
	Facility of manual ventilation in case of electricity and battery failure	Yes
	Facility of gas and agent delivery in case of electricity and battery failure	Yes
	All the components likes anesthesia machine, integrated ventilator, vaporizer from the same manufacturer	Yes
	The workstation should have one or more drawers to keep accessories, good mobility, anti-static caster wheels with locking facility & conveniently placed handles for easy movement of the machine	Yes
	All software updates should be provided free of cost during Warranty period	Yes
	The anesthesia machine can be used safely up to 1000 gauss line while being used with MRI	No
	Pressure relief	5 to 70 cm H20
CERTIFICATIONS	Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid medical device license (CDSCO License) for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules as amended till date	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
WARRANTY	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	2 Or higher **(year)**

व्यापक रखरखाव / **Comprehensive Maintenance**	
Comprehensive Maintenance (Minimum Percentage)	3 %
Comprehensive Maintenance (Maximum Percentage)	25 %
Warranty of required product	5 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specfication

Additional Specification Parameters - Anaesthesia Workstation (V3) (2 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Parameter 1	Each Anaesthesia Workstation shall be supplied with compatible, calibrated and manufacturer-approved vaporizers.

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Manjula Begum	744104,DHS ANNEXE BUILDING, OLD RBD BUILDING, ATLANTA POINT, PORT BLAIR, SOUTH ANDAMAN	2	90

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. **Generic**

   Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

   Anaesthesia Department, G.B.Pant Hospital
   Atlanta Point, Sri Vijaya Puram
   South Andaman
   Andaman & Nicobar Islands
   PIN - 744101
   .

3. **Generic**

   Bidder financial standing: The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

4. **Generic**

   Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

5. **Generic**

   Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

6. **Generic**

   Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

7. **Generic**

   Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

8. **Generic**

   OPTIONAL SITE VISIT:

   1. The Bidder is advised to visit and examine the installation site and its surroundings and obtain for itself on its own responsibility all information that may be necessary for preparing the Bid. The costs of visiting the site shall be borne by the Bidder.

   Cost of visiting the site shall be entirely borne by the bidder.

   2. The Bidder representative shall be allowed entry upon consignee premises for such visits, only upon

the express conditions that the Bidder will release and indemnify the Buyer and Consignee against all liabilities arising out of such visit including death or injury, loss or damage to property, and any other loss, damage, costs, and expenses incurred as a result of such visit.

3. The Bidder shall not be entitled to hold any claim against Buyer for noncompliance due to lack of any kind of pre-requisite information as it is the sole responsibility of the Bidder to obtain all the necessary information with regard to site, surrounding, working conditions, weather etc. on its own before submission of the bid.

9. **Generic**

Scope of supply includes Training: Number of employees to be trained

2
, Place for Training
Anaesthesia Department, G.B.Pant Hospital, Sri Vijaya Puram
and Duration of training
2
days.

10. **Generic**

The successful bidder has to supply all essential accessories required for the successful installation and commissioning of the goods supplied. Besides standard accessories as per normal industry practice, following accessories must be part of supply and cost should be included in bid price:

Cost included in the bid price.
.

11. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

12. **Turnover**

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

13. **Turnover**

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

14. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

15. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free

Telephone No. for Service Support.

16. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

17. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

18. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

19. **Warranty**

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 5 years for the offered products at the rate not more than 10 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Annually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of

10
(Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 2% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

20. **Warranty**

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

21. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 15 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG).Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imberse the cost of such service / rectification to the Buyer.

22. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

DIRECTOR ANIIMS
payable at
SRI VIJAYA PURAM
.
Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy

to the Buyer within 5 days of Bid End date / Bid Opening date.

23. **Forms of EMD and PBG**

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

DIRECTOR ANIIMS
. The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

24. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

DIRECTOR ANIIMS
payable at
SRI VIJAYA PURAM
. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

25. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

DIRECTOR ANIIMS
A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

26. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

1. The Bidder or its OEM must have supplied atleast 03 Nos same or similar category products to any Central/ State Govt. Organization/ PSU in the last 03 financial years. Documentary Evid ence should be submitted along with the Bid.

2. The bidder is required to upload all relevant certificates such as BIS license, type test certifi cate, approval certificates and other certificates as prescribed in the Product Specification giv en in the Bid Document.

3. The Successful Bidder should provide 05 Years of Warranty and 05 Years of CAMC (after exp iry of the warranty period CAMC should be executed). An undertaken should be provided by th e Bidder.

4. Certifications and Regulatory Compliance:

The offered Anaesthesia Workstation shall comply with the applicable provisions of the Drugs and Cosmetics Act, 1940 and Medical Devices Rules, 2017, as amended. The bidder shall submi t valid Indian manufacturing/import licence or other applicable regulatory authorisation docum ents for the offered make, model and configuration, including the supplied vaporizers, wherev er separately applicable.

5. The equipment shall comply with ISO 80601-2-13:2022, including applicable amendments/cu rrent edition, along with relevant general electrical safety and electromagnetic compatibility r equirements, including IEC 60601-1 and IEC 60601-1-2, as applicable. Relevant conformity doc uments, test reports and manufacturer's technical literature shall be submitted. BIS certificati on shall be furnished wherever mandatory under applicable Government orders.

.

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid.All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM.If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents (other than EMD and signed Integrity Pact - which can be submitted within 5 days of bid opening) in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six

months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
3.GST shall be included in the CMC Charges quoted.
4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 15 years can be chosen for CMC charges beyond warranty period.
7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.
8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period– Percentage to be indicated- A1
CMC charges for 2nd year after warranty period– Percentage to be indicated- A2
CMC charges for 3rd year after warranty period – Percentage to be indicated- A3
CMC charges for 4th year after warranty period – Percentage to be indicated- A4
CMC charges for 5th year after warranty period – Percentage to be indicated- A5

Similarly, A6 to A15 are to be indicated for 6th to 15th year of CMC if applicable.
8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.
8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below includingCMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.
(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:
Total Cost for evaluation=
C+C*{(A1/100)/(1.10^n)+(A2/100)/(1.10^n+1)+(A3/100)/(1.10^n+2)+(A4/100)/(1.10^n+3)+ (A5/100)/(1.10^n+4)} and so on
C – Cost for equipment quoted and n shall be number of years of product warranty specified.
If 2 year warranty specified, n shall be2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5shall depend on how many years CMC selected. For3 yearCMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.
8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.
8.5.The CMC charges shall be offered within range of 3 to up to 50% of cost of equipment as defined by buyer.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.
10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.
11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall

be matched individually.
12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.
13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---