बिड संख्या/Bid Number: GEM/2026/B/8243727
दिनांक /Dated: 06-11-2026

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	16-11-2026 13:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	17-11-2026 13:00:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Defence
विभाग का नाम/Department Name	Department Of Defence Research & Development
संगठन का नाम/Organisation Name	Office Of Dg (sss)
कार्यालय का नाम/Office Name	**********
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :yksharma@dipas.drdo.in Buyer Email id: rssuman.dipas@gov.in
कुल मात्रा/Total Quantity	1
वस्तु श्रेणी /Item Category	High quality water purification system (Q2)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	OEM Authorization Certificate,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3

बिड विवरण/Bid Details	
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	3
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	3 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य / Estimated Bid Value	2419767
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाईजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	72593

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%)/ePBG Percentage(%)	3.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	62

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be

in favour of Beneficiary, wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

लाभार्थी /Beneficiary :
DIRECTOR DIPAS
DIPAS, LUCKNOW ROAD, TIMARPUR, DELHI-110054
(Director Dipas)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में है / Purchase Preference to MII sellers availabele upto price within L1+X%	20
मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference	50
सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)	Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time)

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% / Purchase Preference to MSE OEMs/ Service Provider available upto price within L1+X%	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the

Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.
OM_No.1_4_2021_PPD_dated_18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Orders/Notifications issued by concerned Ministry .Benefits of MSE will be allowed only if seller/service provider is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

2. **Purchase preference to Micro and Small Enterprises (MSEs):** Purchase preference will be given to MSEs having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted in case of product bids, whereas in case of services the eligibility is automatically validated. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM_No.1_4_2021_PPD_dated_18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

4. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

i. If number of technically qualified bidders are only 2 or 3.
ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

High Quality Water Purification System (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Performance Parameters	**Purpose**	Lab applications for HPLC/Ion Chromatography / LCMS/Titration
	Prefilteration RO Unit supplied	Yes Or higher
	In put quality in TDS for prefileteration unit	Prefilteration RO Not provided
	Out put quality from prefilteration RO Unit in TDS	300<
	Out put flow rate for pre filteration unit	Minimum 50 litre per hour to be provided, NA if pre filteration unit not provided
	In case pre filteration unit provided capacity of water storage tank from pre filteration unutb	Minimum 50 litre
	Type of water produced by system	Type1 and Type 2
	Resistivity of water out put	10-15 mega ohms .cms or better @25 degree celsius (For Type I)
	Toc (ppb)	<30 (Type 2)
	Display provided	LED
	Active carbon beads Electro Deionization Module (EDI)	Yes Or higher
	Provision of UV in storage tank	Yes Or higher
Miscellaneous Parameters	**For easy understanding of main operations Built in quick reference guide in system**	Yes Or higher
	Number of installations of the Machine machines in Central /State/PSU Govt Hospitals(Note: Seller should supply a performace certificate of the device to the buyer if demanded after placement of order)	More than 3

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	**********	**********NORTH DELHI	1	60

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. **OEM**

   IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

2. **Buyer Added Bid Specific ATC**

   Buyer Added text based ATC clauses

Technical Specification of Type-I and Type-II Water Purification

Part 1: High Quality Pre-Filter: Pre-filter for attaching before main water filtration units to counter the feed water contamination. Pre-filter should be manufactured by same equipment OEM with following specifications: high quality pre-filter with low pressure switch cuts off system which can able to take care high TDS up to 5000 ppm.

Apart from pre-filters two separate model should be there for Type-I and Type II systems.

Part 2: Lab Grade Water Purification System(Type II): Feed water acceptance capability specifications for main unit:

(a) Conductivity: <0.2µS/cm at 25°C (water output)

(b) Resistivity: Typically 5-15 mega Ohms

(c) flow rate – 3 Ltrs/hr

(d) TOC: < 30 ppb

(e) Fouling Index(SDI): upto 12.

(f) Free Chlorine: upto 3 ppm.

(g) Temperature-5-35ºC.

Type II water should meet

1) System should have facility to control remotely with the help of software interface. All parameters like pressure, current, water rejection %, water product quality should be monitored in desktop or laptop.

2) Must have automatic electrodeionization module.

3) EDI must have carbon beads at cathode which doesn’t require extra pre softening cartridges for better water quality or IR-control with proactive gas filling

under argon.

4) No of conductivity cell in Type-II to ensure the Input & Output water quality-3 Nos.

5) Inside Type-II system inbuilt pre-treatment Cartridge: 0.5 Micron filter, anti-scaling compounds and RFID tag must be in this system.

6) Pre-treatment pack contains silver-impregnated activated carbon which prevents the proliferation of bacteria present in tap water; anti scaling compounds must eliminate hardness and protect the RO membrane against oxidation, scaling and plugging.

7) RFID Tag for traceability of per-treatment cartridge and every litre throughput water volume should be seen on computer when connected through LAN. The record for date of cartridge installation must be available on computer.

8) RO reject water recovery upto 50% before RO conductivity cell to know the % rejection of RO

9) In Type II system:- The following parameter must be visible on system LCD/LED display:

a) Tap water feed conductivity.

b) RO feed water conductivity.

c) RO water temperature.

d) RO pump pressure.

e) RO % rejection value.

f) RO permeate water conductivity.

g) Tank level

10) 3 way solenoid valve- RO permeate is diverted to drain until the quality meets expectations. There must be recovery loop in II system to save wastage of rejected water.

11) Warranty in years: **Minimum 5Years**

Storage Reservoir: Tank capacity: 50 Liters

Part 3: Ultra-Pure Water (Type I) should meet:

Typical Water Delivery flow Rate: 2L/min

Ultrapure Water Resistivity (MΩ.cm at 25ºC) : 18.2

Conductivity: ≤ 0.055µs/cm at 25ºC

water particulates: No particles with a size > 0.22 mm

TOC(ppb): ≤5 (TOC indicator should be there)

Microorganisms (cfu/mL): End filter<0.01

Pyrogen Levels (EU/mL): <0.001

RNase Level(pg/mL): <1

DNase Level (pg/mL): <5

Protease (ug/mL): <0.15

Recirculation loop:1 No

1) The water system has a built-in UV lamp with emission at 185 nm and 254n m wavelength. The lamp will operate for 2 years without replacement in orde r to reduce operating costs and the maintenance procedures.

2) Two numbers of polishing cartridge of mixed bed resins should be there in Ty pe-I system .

3) The system will have a choice of minimum 6 point-of use final filter on the w ater dispenser to support different applications. Submit details of all minimu m six.

4) The water Point-of-Delivery can be installed on a wall-mounted support for ful l flexibility of installation. The wall mounted configuration where the HMI and the Point of Delivery are fixed on the wall to a distance up to 3 meters. POD dispensing experience should be the same whatever the installation of the w ater purification system.

5) At the Point-of-Delivery, basic information on water quality or any alert / alar m information should be shared.

6) 7” large or bigger touchscreen with bright color LED/LCD display compatible with gloves and wet or dry hands.

7) Only one single inline photo-oxidation UV lamp allows both to reduce organic s in ultra-pure water and to measure TOC level.

8) Inline TOC indicator is being measure at the point of use and displayed on th e touchscreen after each dispense (Real time TOC monitoring).

9) UV lamp should be made up of mercury free or low pressure Mercury.

10) The water purification system is able to report the resistivity and conductivity in both temperature compensated and non-temperature compe nsated formats. The resistivity meters are able to verified and calibrated on-site.

11) The system can be connected to the local network via an Ethernet port and has a USB key port.

12) Must have possibility for connecting extra dispenser for ICP-MS application.

13) Water dispensing reports record all with dispensing parameters (water quali ty and volume, data and time).

14) The volumetric water dispensing to deliver a specific water volume from 10 0ml to 25L automatically (with 100ml incremental). A quick volumetric dispe nsing is also accessible from a pre-set menu of customizable water volumes.

15) Installation and warranty should be taken care by OEM service engineer.

16) CE certificates for Type 1 & Type II system specific for make/model need to be submitted along with tender.

17) ISO Certificate: 13485 & 9001 or latest

18) Must have possibility for Type I water with boron <0.50ppb.

19) Warranty in years: **Minimum 5Years**

20) Demonstration of equipment and training should be provided after completi

ng supplies before acceptance.

21) Should be availability of toll free facility for technical support maintained by OEM or authorized agencies.

22) Should be provided number of installations of the machines in Central/State /PSU Govt Hospitals.

.

3. **Buyer Added Bid Specific ATC**

   Buyer uploaded ATC document Click here to view the file.

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid.All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM.If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents (other than EMD and signed Integrity Pact - which can be submitted within 5 days of bid opening) in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of

this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---