बिड संख्या/Bid Number: GEM/2026/B/8244082

दिनांक /Dated: 05-11-2026

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	26-11-2026 15:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	27-11-2026 15:00:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Uttar Pradesh
विभाग का नाम/Department Name	Medical Health And Family Welfare Department Uttar Pradesh
संगठन का नाम/Organisation Name	N/a
कार्यालय का नाम/Office Name	Lucknow Hq
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :reebha.j@ias.nic.in Buyer Email id: buycon6.upmsc.lko@gembuyer.in
कुल मात्रा/Total Quantity	1
वस्तु श्रेणी /Item Category	Echo Cardiography Scanning Machine (V2) (Q2)
पिछले 3 वित्तीय वर्षों का बिडर का न्यूनतम औसत वार्षिक वित्तीय टर्नओवर/Minimum Average Annual Turnover of the bidder (For the last 3 Financial Years)	18.6 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
टर्नओवर पात्रता मानदंड/Turnover eligibility criteria	To be verified by the buyer at the time of technical evaluation
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No

बिड विवरण/Bid Details	
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	No
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	10 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	50% Lowest Priced Technically Qualified Bidders
व्यापक रखरखाव शुल्क आवश्यक / Comprehensive Maintenance Charges Required	Yes
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
वित्तीय दस्तावेज की आवश्यकता है / Financial Document Required	Yes
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाईजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	124000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%)/ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	38

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

लाभार्थी /Beneficiary :
Managing Director
Lucknow
(Managing Director, Upmscl)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% / Purchase Preference to MSE OEMs/ Service Provider available upto price within L1+X%	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. **Purchase preference to Micro and Small Enterprises (MSEs):** Purchase preference will be given to MSEs from the State of Bid Inviting Authority having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry/ State of Bid Inviting Authority. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Orders/Notifications issued by concerned Ministry. Benefits of MSE will be allowed only if seller/service provider is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
4. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
5. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

6. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

एक्सेल में अपलोड किए जाने की आवश्यकता /Excel Upload Required :

Accessories Cost if Any - 1793869561.xlsx

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
13-11-2026 14:00:00	3rd Floor UPMSCL, SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow

Echo Cardiography Scanning Machine (V2) (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	Echo Cardiography Scanning Machine
Types of Probes & Features	**Cardiac and Vascular Probes Configuration**	TTE Matrix Array
	TTE(Transthoracic) Probe Type	Adult, Pediatric, Neonatal
	TEE Adult Probe biplane and Multiplane frequency, in MHz	2 to 7.5 MHz (±2), Not Applicable
	TEE Pediatric Probe Required	Yes, No
	TEE Pediatric Probe frequency, in MHz	3 to 8 MHz (±2), Not Applicable
	Linear Probe for linear/ vascular applications required	Yes, No
	Adult/Pediatric Linear probe frequency in MHz	5 to 10MHz (±2)
	No of Adult TTE Probes Required	1
	No of Pediatric TTE Probes Required	1
	No of Neonatal TTE Probes Required	1, NA
	No of Adult TEE Probes Required	1, NA
	No of Pediatric TEE Probes Required	1, NA
	No of Pediatric 3D/4D TEE Probes Required	1, NA
	No of Neonatal TEE Probes Required	NA
	No of Adult/pediatric Linear Probes supplied	NA

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Dedicated 3D/4D volume imaging for cardiology TTE ad TEE	Yes, No
	Live 3 D imaging in color doppler system	Yes, No
	Colour Doppler system with all application packages ,Quad loop for serial studies with high frame rate revies, harmonic imaging capability in all modes and integrated stress echo package digital storage and retrieval	Yes
	Provision of Tissue Colorization (B-Colour) for improved contrast resolution	Yes
	Application software	Adult, Pediatric, Neonatal, Vascular applications and Transesophagal applications
	Cine Loop Features	Yes
	Frame grabber facility for post analysis	Yes
	Facility of Various maps for pre and post processing	Yes
	ECG trigger facility	Yes
	User defined Preset	Yes
	Dedicated integrated dynamic stress echo package for flexible user defined protocols with stacked sub loops facility and contrast stress protocol usable for stress echocardiography	Yes, No
	Number of transducer/probe ports	4
Scan Modes and Performance Parameters	**Cardiac 3-D freehand Option**	Yes, No
	Cardiac 3D/4D automatic Option	Yes, No
	Contrast Harmonic/Agent Imaging with angio mode	Yes, No
	Color Doppler 3-D/4-D option	Yes, No
	Power Dopple 3-D/4-D option	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Contrast-enhanced ultrasound (CEUS) option	Yes
	Tissue synchronization, tissue velocity	Yes
	Cardiac Shear wave elastography	Yes, No
	Cardiac compression elastography	Yes, No
	Provision for higher sensitivity for low frequency doppler in all probes	Yes
	Latest generation Electronic Phased/sector array Colour Doppler system	Yes
	Number of Channels in color doppler system	5000000 or more (if channels are being designated as digital channels)
	System should be capable of generating real time live 3- D images	Yes, No, Yes with MV No Analysis
Image display and processing	**DICOM 3 Compliant**	Yes
	Automated B-mode (2-D) image	Yes
	Automated CDI image optimization	Yes
	Automated PW Doppler image optimization	Yes
	Touch Screen	Yes
	Monitor display size, cm (in)	21 Inch(±1) Or higher
	Split screen	Twin View and Live Quad Screen
	Single/dual monitors	Single with optional second console
	DIGITAL IMAGE STORAGE	Yes
	CINE LOOP PLAYBACK	Yes
	Speckle-tracking strain and strain rate	Yes, No
	Analysis and calculation packages	2-D wall motion tracking, 3-D wall motion tracking, Cardiac, Mitral valve analysis, User-programmable calculations, No
	Customizable presets	Yes
	On-screen annotation	Yes
	Flash Memory	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	USB connector	Yes
	Type of printer provided with the echo cardio machine	Themal Printer, Color Laser Printer
	Provision of PC with echo cardio gram machine	Yes
	Type of processor in PC	i7 or Latest
	Conectivity between Echo cardiogram and PC	Through IMS/Equivalent Software
	PC with all software inclusive interfaced with echocardiograph machine	Yes
	Inkjet printer provided with computer for reporting	Yes
	Glossy Colour Print Paper supplied	500 sheets, 1000 sheets, Not provided
	ECG Cable	2, NA
	Echo cardiography sytem to be DICOM Ready	Yes
	Dicom Interfacability to HIS	Yes
	Dicom Interfacability Dicom to PACS	Yes
	Interfacability Dicom to RIS	Yes
	Echo Cardiography sytem supplied should be upgradable to next generation system on site	Yes
Power Requriments	Power bakup	UPS
	Back up time in minutes	30, 60, 90, 120 Or higher **(minute)**
CERTIFICATION AND REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for product as per Medical Device Rules (MDR) 2017 as amended till date	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
MISCELLANEOUS PARAMETERS	Availability of toll free facility for technical support maintened by OEM or authorized agencies	Yes
	Supplier to perform installation, safety and operation checks before handover	Yes
	Training of users in operation and basic maintenance shall be provided	Yes
	Contact details of manufacturer, supplier and local service agent to be provided	Yes
	Certificate of calibration from the manufacturer	Yes

व्यापक रखरखाव / **Comprehensive Maintenance**	
Comprehensive Maintenance (Minimum Percentage)	3 %
Comprehensive Maintenance (Maximum Percentage)	25 %
Warranty of required product	3 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specfication

Additional Specification Parameters - Echo Cardiography Scanning Machine (V2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Analysis and calculation packages: 2-D wall motion tracking, 3-D wall motion tracking, Cardiac, Mitral valve analysis, User-programmable calculations, No	Kindly Read As Analysis and calculation packages: 2-D wall motion tracking, Cardiac, Mitral valve analysis, User-programmable calculations
TTE(Transthoracic) Probe Type: Adult, Pediatric, Neonatal	Kindly Read as TTE(Transthoracic) Probe Type: Adult, Pediatric

Specification Parameter Name	Bid Requirement (Allowed Values)
Application software: Adult, Pediatric, Neonatal, Vascular applications and Transesophagal applications	Kindly read as Application software: Adult, Pediatric
Quality Certificate	1- EU CE from 4 digit Notified Body/USFDA/BIS Certificate 2- CDSCO Licence 3- ISO 13485 Certificate

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Aarti Pandey	226003,Room No. 714, Vikas Bhawan, Janpath, Lucknow-226001	1	60

Special terms and conditions-Version:1 effective from 15-04-2024 for category Echo Cardiography Scanning Machine (V2)

1. 1. *A*ll Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
   2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
   3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
   4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
   5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
   6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
   7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration

of service requests also to be provided buyer/consignee at the time of supplies.

8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transhipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. **Generic**

   OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. **Generic**

   Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

   Consignees are various locations in different districts of UP. However actual consignee will be shared at lat er stage after the finalization of said tender.
   Consignees are various locations in different districts of UP. However actual consignee will be shared at lat

er stage after the finalization of said tender.
Consignees are various locations in different districts of UP. However actual consignee will be shared at lat er stage after the finalization of said tender.
Consignees are various locations in different districts of UP. However actual consignee will be shared at lat er stage after the finalization of said tender.
Consignees are various locations in different districts of UP. However actual consignee will be shared at lat er stage after the finalization of said tender.
.

3. **Scope of Supply**

   Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

4. **Turnover**

   Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

5. **Certificates**

   Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

6. **Forms of EMD and PBG**

   Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

   Managing Director, Medical Supplies Corporation Limited, Lucknow
   . The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

7. **Forms of EMD and PBG**

   Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name

   UP MEDICAL SUPPLIES CORPORATION LIMITED
   Account No.
   39366886265
   IFSC Code
   SBIN0012732
   Bank Name
   STATE BANK OF INDIA
   Branch address
   Arjunganj Lucknow
   .
   Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

8. **Buyer Added Bid Specific ATC**

   Buyer Added text based ATC clauses

   1- The physical demonstration of the said item will be held, details of which will be shared at later stage wi th technically qualified bidders through email. Therefore bidders are requested to mention their email id a

nd contact no. in the bid document.

2- It may be noted that transportation charges or any other charges for physical demonstration will not be borne by UPMSCL. It is the responsibility of respective bidders.

3- EMD exemption applicable as per the MSE Purchase Policy 2020.

4-Successful bidder shall upload all documents related to delivery, installation, commissioning, warranty, a nd any other documents required for processing of payment through the EMMS Portal. Submission of the r equisite documents on the EMMS Portal shall be mandatory for the release of payment

.

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid.All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM.If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents (other than EMD and signed Integrity Pact - which can be submitted within 5 days of bid opening) in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such

representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
3.GST shall be included in the CMC Charges quoted.
4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 15 years can be chosen for CMC charges beyond warranty period.
7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.
8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period– Percentage to be indicated- A1
CMC charges for 2nd year after warranty period– Percentage to be indicated- A2
CMC charges for 3rd year after warranty period – Percentage to be indicated- A3
CMC charges for 4th year after warranty period – Percentage to be indicated- A4
CMC charges for 5th year after warranty period – Percentage to be indicated- A5

Similarly, A6 to A15 are to be indicated for 6th to 15th year of CMC if applicable.
8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.
8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below includingCMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.
(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:
Total Cost for evaluation=
C+C*{(A1/100)/(1.10^n)+(A2/100)/(1.10^n+1)+(A3/100)/(1.10^n+2)+(A4/100)/(1.10^n+3)+(A5/100)/(1.10^n+4)} and so on
C – Cost for equipment quoted and n shall be number of years of product warranty specified.
If 2 year warranty specified, n shall be2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5shall depend on how many years CMC selected. For3 yearCMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.
8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.
8.5.The CMC charges shall be offered within range of 3 to up to 50% of cost of equipment as defined by buyer.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance

guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.
10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.
11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall be matched individually.
12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.
13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---